

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

This is to inform you that B. Braun Medical Inc. (BBMI) is issuing a pharmaceutical recall for product P9402, KCL CONC 2MEQ/ML INJ USP 250/ML, due to the potential for fluid leakage. BBMI has identified through complaints the potential for fluid leakage originating from pinholes within the blue label characters on the EXCEL bag, specifically within the dotted characters on the label.

To date, there have been no reports of serious injury or death associated with this issue. Unidentified leakage could lead to contamination with microorganisms causing the potential for bloodstream infection from product usage which could become a life-threatening event.

Product Catalog Number	NDC Number	Product Description	Lot Number	Distribution Range	Expiration Date	Region Distributed
P9402	0264-1944-20	KCL CONC 2MEQ/ML INJ USP 250 ML	J2S007	1/30/23 - 1/30/23	12/31/2024	US
P9402	0264-1944-20	KCL CONC 2MEQ/ML INJ USP 250 ML	J3A115	2/28/23 - 4/10/23	1/31/2025	US