

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

Description	Lot #	UPC
ADM 20 DRP INTEGRAT CLAVE 3011	5125297	
CHEMOLOCK 30 SET DRIP 3011 50	5125298	
CHEMOLOCK ADMIN SET CL3028T 50	5125199	88770999962
CHEMOLOCK EXT SET CL4151 50	5142912	88770998389
ONCOKIT W/VNT CAP CL3946ICU50	5142959	88770908029
ONCOLOGY KT W/5 CAP 3538 50	5125294	
SPIN SPIR CLS M LUER 2000S 100	5125299	
SPIN SPIR MALE LUER 2000S-C 50	5125303	84061902661
SPIROS MALE CONNECT+CAP CS50	5142982	30409201301
	5135750	
SPIRS MALE LR 2000SC-10 10UN	5142969	

ICU Medical is recalling the above items/lots due to the potential for certain lots of Spiros to exhibit small amounts of leaks due a molding defect. This recall is to the retail level. Affected product started shipping January 12, 2021.

NOTE: Additional affected products/lots are not listed here. For additional information:

ICU Medical Contact	Contact Information	Areas of Support
Global Complaint Management	1-844-654-7780 or ProductComplaintsPP@icumed.com	To report adverse events or product complaints
Customer Service	1-866-829-9025, option 8 or customerservice@icumed.com (M-F, 8am-6pm CT)	Additional information or assistance