

The Board of Pharmacy has received notice of the following product recall:

Baxter Healthcare Corporation is expanding the product recall of EXACTAMIX Empty EVA Bags issued on September 12, 2019, **to include two additional lots listed in bold below** due to the potential ability to leak once used for compounding. The two additional lots were distributed between 2/13/2018 and 1/7/2019 in the United States.

<b>Product Code</b>	<b>Product Description</b>	<b>Lot Numbers</b>	<b>Expiration Dates</b>	<b>UDI#</b>
H938738	EXACTAMIX EVA Bag - 500mL	60137163	6/30/2021	00085412477220
H938739	EXACTAMIX EVA Bag - 1000mL	60072116	6/30/2020	00085412477237

The Board of Pharmacy strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.