

The Board of Pharmacy has received notice of the following product recall. The Board strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.

Description	Lot #	Exp Date	NDC	UPC
DUODERM CGF DRESS 4X4 CONV 5	9L00556	11/01/2024	00003187660	76845510697
	9G04308	07/01/2024		
	9B02984Y	02/01/2024		
	9A04124Y	01/01/2024		
DUODERM X/T 6X6 187957 CONV 10	9L02456	11/01/2024	00003187957	76845510692
	9H04865	08/01/2024		

ConvaTec is recalling the above items/lots due to an increasing trend of complaints related to open seals, tears and rips in the primary packaging. This recall is to the healthcare professional level. Affected product started shipping March 15, 2019.