

The Board of Pharmacy has received notice of the following product recall:

LEVEL OF NOTIFICATION: Retail pharmacy

SUPPLIER: BD Medical

Description	Lot #	Exp Date	NDC	UPC
SYR POSIF SAL SHRT 10ML BD 480	9142881	04/30/2022	08290306547	38290306547

BD Medical is voluntarily recalling the above item/lot as BD has become aware that a limited number of syringes labeled "Posiflush Experimental Product" and "Not for Human Use" being mixed with standard BD Posiflush Pre-Filled Normal Saline Syringes. This recall is to the retail pharmacy level. Affected product started shipping June 2019.

The Board of Pharmacy strongly encourages pharmacies to immediately review their quality assurance and recall policies and procedures to determine if any corrective action is required.